

EC Declaration of Conformity

Manufacturer: Xiamen AmonMed Biotechnology Co.,Ltd.

Address: Unit 503, 120 Xinyuan Road, Haicang District, Xiamen, Fujian, China

EC Representative: SUNGO Europe B.V.

Address: Olympisch Stadion 24,1076DE Amsterdam, Netherlands

Product Name: COVID-19 Antigen Rapid Test Kit (Colloidal Gold)

Product specification and Reference No
For use with saliva specimen:
1 test/kit: CG01Ag-01S-ST
5 tests/kit: CG01Ag-05S-ST
25 tests/kit: CG01Ag-25S-ST

Classification: Self-test IVD

Conformity-Assessment Procedure: Directive (98/79/EC) on In Vitro Diagnostic Medical Device, Annex III section 6

We herewith declare that the above mentioned products meet the provisions of In Vitro Diagnostic Directive (98/79/EC). All supporting documentation is retained under the premises of the manufacturer. We have sole responsibility for issuing the Declaration of Conformity.

Applied standards:	EN ISO 18113-4:2011	EN ISO 23640:2015
EN ISO 13485:2016	EN ISO 18113-1:2011	EN ISO 15223-1:2016
EN ISO 14971:2019	EN ISO 17511:2003	EN 13612:2002
ISO/TR 24971-2020	EN 13532:2002	EN 13641:2002
EN 62366:2015		
MDCG 2021-21		

Notified Body: Polish Center for Testing and Certification, 469 Putawska Street, 02-944 Warsaw
Identification number: 1434
CE certificate No.

Signature:

Name and Position: Fenfang Xie/Management Representative

First Issue Date:

2021.10.15

Issue Date:

2021.10.15

